



International Accreditation Japan

Information on Accredited Reference Material Producer

Date of the update of the information : 2026-04-01

Accreditation Identification: ASNITE 0001 RMP

Name of Reference Material Producer : National Metrology Institute of Japan,
National Institute of Advanced Industrial Science and Technology

Location of Reference Material Producer : 1-1-1 Umezono, Tsukuba-shi, Ibaraki
305-8563, JAPAN

Name of Legal Entity: National Institute of Advanced Industrial Science and Technology

Conformance Accreditation Standard: ISO 17034:2016

Expiry Date of Accreditation : 2029-10-31

Subcategory	Measurand		Measurement Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Standard gases	high purity nitrogen monoxide (NO)		0.99 mol/mol to 0.99993 mol/mol	1.0 % to 0.01 % (relative)	• Subtraction method	2024-11-01
	impurities in NO	NO ₂	10 µmol/mol to 10000 µmol/mol	10 % to 2.5 % (relative)	• FT-IR	
		N ₂	11 µmol/mol to 5000 µmol/mol	100 % to 2.5 % (relative)	• GC-TCD	
		O ₂	11 µmol/mol to 5000 µmol/mol	100 % to 2.5 % (relative)	• GC-TCD	
		N ₂ O	7.5 µmol/mol to 11000 µmol/mol	10 % to 0.5 % (relative)	• FT-IR • GC-TCD	
		CH ₄	2 µmol/mol to 11000 µmol/mol	100 % to 2.5 % (relative)	• FT-IR • GC-FID	
		C ₃ H ₈	2 µmol/mol to 11000 µmol/mol	100 % to 2.5 % (relative)	• GC-FID	
		H ₂ O	21 µmol/mol to 100 µmol/mol	100 % to 0.5% (relative)	• FT-IR	
		CO ₂	10 µmol/mol to 100 µmol/mol	100 % to 0.5% (relative)	• FT-IR	
	high purity sulfur dioxide (SO ₂)		0.99 mol/mol to 0.99997 mol/mol	1.0 % to 0.01 % (relative)	• Subtraction method	
	impurities in SO ₂	CO ₂	1 µmol/mol to 15000 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD • FT-IR	
		N ₂	1 µmol/mol to 15000 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD	
		O ₂	1 µmol/mol to 15000 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD	
		CH ₄	0.09 µmol/mol to 11000 µmol/mol	100 % to 0.5 % (relative)	• GC-FID	
		C ₃ H ₈	0.04 µmol/mol to 11000 µmol/mol	100 % to 0.5 % (relative)	• GC-FID	
		H ₂ O	24 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• FT-IR	
	high purity methane (CH ₄)		0.99 mol/mol to 0.999995 mol/mol	1 mmol/mol to 0.0005 mmol/mol	• Subtracting method	
	impurities in CH ₄	N ₂	0.1 µmol/mol to 100 µmol/mol	80 % to 2 % (relative)	• GC-PID • GC-TCD	
		O ₂	0.1 µmol/mol to 100 µmol/mol	60 % to 2 % (relative)	• GC-PID • GC-TCD	
		Ar	0.1 µmol/mol to 100 µmol/mol	40 % to 2 % (relative)	• GC-PID • GC-TCD	
		CO	0.04 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-PID • GC-TCD	
CO ₂		0.04 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-PID • GC-TCD		
C ₂ H ₆		0.02 µmol/mol to 100 µmol/mol	100 % to 2 % (relative)	• GC-FID		
H ₂		0.07 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-PID • GC-TCD		
hexane		0.02 µmol/mol to 20 µmol/mol	100 % to 0.6 % (relative)	• GC-FID		
H ₂ O		0.1 µmol/mol to 130 µmol/mol	70 % to 5 % (relative)	• Dew point measuring method		

Subcategory	Measurand		Measurement Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Standard gases	high purity propane (C ₃ H ₈)		0.99 mol/mol to 0.999998 mol/mol	1 mmol/mol to 0.001 mmol/mol	• Subtracting method	2024-11-01
	impurities in C ₃ H ₈	N ₂	3 µmol/mol to 100 µmol/mol	80 % to 2 % (relative)	• GC-TCD	
		O ₂	0.1 µmol/mol to 100 µmol/mol	60 % to 2 % (relative)	• GC-TCD	
		Ar	0.1 µmol/mol to 100 µmol/mol	40 % to 2 % (relative)	• GC-TCD	
		CO ₂	0.1 µmol/mol to 100 µmol/mol	50 % to 2 % (relative)	• GC-TCD	
		CH ₄	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-FID	
		C ₂ H ₆	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-FID	
		propylene	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-FID	
		butane	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-FID	
		isobutane	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-FID	
		H ₂ O	10 µmol/mol to 1000 µmol/mol	70 % to 20 % (relative)	• Dew point measuring method	
	high purity carbon dioxide (CO ₂)		0.99 mol/mol to 0.999998 mol/mol	1 mmol/mol to 0.002 mmol/mol	• Subtracting method	
	impurities in CO ₂	N ₂	0.1 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD	
		O ₂	0.1 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD	
		H ₂	0.8 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD	
		He	0.8 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD	
		CH ₄	0.004 µmol/mol to 1 µmol/mol	100 % to 1 % (relative)	• GC-FID	
		C ₃ H ₈	0.004 µmol/mol to 1 µmol/mol	100 % to 1 % (relative)	• GC-FID	
		CO	0.05 µmol/mol to 1 µmol/mol	100 % to 0.5 % (relative)	• GC-FID	
		H ₂ O	0.9 µmol/mol to 130 µmol/mol	100 % to 30 % (relative)	• Capacitance-type moisture analyzer	
	high purity carbon monoxide (CO)		0.99 mol/mol to 0.99993 mol/mol	1 mmol/mol to 0.02 mmol/mol	• Subtracting method	
impurities in CO	N ₂	1.5 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD		
	O ₂	2.1 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD		
	H ₂	0.9 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD		
	He	0.4 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD		
	CH ₄	1.5 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD		
	CO ₂	0.3 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• GC-TCD		
	H ₂ O	0.36 µmol/mol to 100 µmol/mol	100 % to 0.5 % (relative)	• Quartz-crystal oscillator sample cell		

Subcategory	Measurand		Measurement Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Standard gases	high purity oxygen (O ₂)		0.99 mol/mol to 1 mol/mol	1 mmol/mol to 0.0005 mmol/mol	• Subtracting method • Magnetopneumatic oxygen analyzer	2024-11-01
	impurities in O ₂	Ar	1 μmol/mol to 100 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		N ₂	1 μmol/mol to 100 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		CH ₄	0.05 μmol/mol to 1 μmol/mol	30 % to 5 % (relative)	• FT-IR	
		CO	0.06 μmol/mol to 1 μmol/mol	30 % to 5 % (relative)	• FT-IR	
		CO ₂	0.05 μmol/mol to 1 μmol/mol	30 % to 5 % (relative)	• FT-IR	
		N ₂ O	0.05 μmol/mol to 1 μmol/mol	30 % to 5 % (relative)	• FT-IR	
		H ₂ O	0.5 μmol/mol to 130 μmol/mol	70 % to 30 % (relative)	• Dew point measuring method	
	high purity vinyl chloride		0.99 mol/mol to 0.99999 mol/mol	5 mmol/mol to 0.01 mmol/mol	• Subtracting method	
	impurities in vinyl chloride	N ₂	1 μmol/mol to 100 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		O ₂	1 μmol/mol to 100 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		Ar	1 μmol/mol to 100 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		CO ₂	1 μmol/mol to 100 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		methyl chloride	1 μmol/mol to 200 μmol/mol	30 % to 2 % (relative)	• GC-FID	
		ethyl chloride	1 μmol/mol to 100 μmol/mol	20 % to 2 % (relative)	• GC-FID	
		H ₂ O	10 μmol/mol to 1000 μmol/mol	70 % to 20 % (relative)	• Dew point measuring method	
	high purity 1,3-butadiene		0.98 mol/mol to 0.99996 mol/mol	20 mmol/mol to 1 mmol/mol	• Subtracting method	
	impurities in 1,3-butadiene	N ₂	5 μmol/mol to 1000 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		O ₂	5 μmol/mol to 1000 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
		Ar	5 μmol/mol to 1000 μmol/mol	30 % to 2 % (relative)	• GC-TCD	
CO ₂		5 μmol/mol to 1000 μmol/mol	30 % to 2 % (relative)	• GC-TCD		
butane		1 μmol/mol to 500 μmol/mol	20 % to 2 % (relative)	• GC-FID		
isobutane		1 μmol/mol to 500 μmol/mol	20 % to 2 % (relative)	• GC-FID		
1-butene		1 μmol/mol to 1000 μmol/mol	20 % to 2 % (relative)	• GC-FID		
<i>trans</i> -2-butene		1 μmol/mol to 7000 μmol/mol	20 % to 2 % (relative)	• GC-FID		
<i>cis</i> -2-butene		1 μmol/mol to 8000 μmol/mol	20 % to 2 % (relative)	• GC-FID		
isobutylene		1 μmol/mol to 1000 μmol/mol	20 % to 2 % (relative)	• GC-FID		
4-vinyl-1- cyclohexene (1,3-butadiene dimer)		1 μmol/mol to 2150 μmol/mol	60 % to 30 % (relative)	• GC-FID		
H ₂ O		10 μmol/mol to 1000 μmol/mol	70 % to 20 % (relative)	• Dew point measuring method		

Subcategory	Measurand	Measurement Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation	
Standard gases	high purity ethane	0.99 mol/mol to 0.99999 mol/mol	1 mmol/mol to 0.001 mmol/mol	• Subtracting method	2024-11-01	
	impurities in ethane	N ₂	0.1 µmol/mol to 100 µmol/mol	80 % to 2 % (relative)		• GC-TCD
		O ₂	0.1 µmol/mol to 100 µmol/mol	60 % to 2 % (relative)		• GC-TCD
		CO ₂	0.1 µmol/mol to 100 µmol/mol	50 % to 2 % (relative)		• GC-TCD
		methane	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		ethylene	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		propane	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		propylene	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		butane	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		H ₂ O	10 µmol/mol to 1000 µmol/mol	70 % to 20 % (relative)		• Dew point measuring method
	high purity isobutane	0.99 mol/mol to 0.99995 mol/mol	2 mmol/mol to 0.005 mmol/mol	• Subtracting method		
	impurities in isobutane	N ₂	1.76 µmol/mol to 100 µmol/mol	100 % to 2 % (relative)		• GC-TCD
		O ₂	5 µmol/mol to 100 µmol/mol	100 % to 2 % (relative)		• GC-TCD
		CO ₂	11 µmol/mol to 100 µmol/mol	100 % to 2 % (relative)		• GC-TCD
		propane	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		butane	0.1 µmol/mol to 200 µmol/mol	30 % to 2 % (relative)		• GC-FID
		isobutene	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		<i>cis</i> -2-butene	0.1 µmol/mol to 500 µmol/mol	30 % to 2 % (relative)		• GC-FID
		<i>trans</i> -2-butene	0.1 µmol/mol to 500 µmol/mol	30 % to 2 % (relative)		• GC-FID
		pentane	3 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		H ₂ O	50 µmol/mol to 3000 µmol/mol	70 % to 10 % (relative)		• Dew point measuring method
	high purity butane	0.99 mol/mol to 0.99995 mol/mol	2 mmol/mol to 0.005 mmol/mol	• Subtracting method		
	impurities in butane	N ₂	1.76 µmol/mol to 100 µmol/mol	100 % to 2 % (relative)		• GC-TCD
		O ₂	1.7 µmol/mol to 100 µmol/mol	100 % to 2 % (relative)		• GC-TCD
		CO ₂	11 µmol/mol to 100 µmol/mol	100 % to 2 % (relative)		• GC-TCD
		propane	0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)		• GC-FID
		isobutane	1 µmol/mol to 200 µmol/mol	30 % to 2 % (relative)		• GC-FID
isobutene		0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-FID		
<i>cis</i> -2-butene		0.1 µmol/mol to 500 µmol/mol	30 % to 2 % (relative)	• GC-FID		
<i>trans</i> -2-butene		0.1 µmol/mol to 500 µmol/mol	30 % to 2 % (relative)	• GC-FID		
pentane		0.1 µmol/mol to 100 µmol/mol	30 % to 2 % (relative)	• GC-FID		
H ₂ O		50 µmol/mol to 3000 µmol/mol	70 % to 10 % (relative)	• Dew point measuring method		

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation	
Standard gases	high purity isopentane	0.99 mol/mol to 1 mol/mol	5 mmol/mol to 0.01 mmol/mol	• Post-column reaction gas chromatography	2024-11-01	
	high purity pentane	0.99 mol/mol to 1 mol/mol	5 mmol/mol to 0.01 mmol/mol	• Post-column reaction gas chromatography		
	nitrogen	0.999 mol/mol to 0.999998 mol/mol	1 mmol/mol to 0.004 mmol/mol	• Subtracting method		
	impurities in nitrogen	O ₂ +Ar	1 μmol/mol to 10 μmol/mol	100 % to 30 % (relative)		• GC-TCD
		carbon dioxide	0.1 μmol/mol to 10 μmol/mol	100 % to 30 % (relative)		• GC-FID
		total hydrocarbons	0.005 μmol/mol to 10 μmol/mol	100 % to 30 % (relative)		• Total hydrocarbon analyzer
		H ₂ O	1.4 μmol/mol to 10 μmol/mol	100 % to 30 % (relative)		• Dew point measuring method
		O ₂ /N ₂	5 μmol/mol to 5 mmol/mol	1 % to 0.1 % (relative)		• GC-TCD
		N ₂ O/N ₂ or N ₂ O/air	0.2 μmol/mol to 0.02 mol/mol	0.2 % to 0.1 % (relative)		• GC-TCD • GC-ECD
		hexane/N ₂	20 μmol/mol to 600 μmol/mol	2 % to 0.3 % (relative)		• GC-FID
		hexane/CH ₄	20 μmol/mol to 600 μmol/mol	2 % to 0.3 % (relative)		• GC-FID
		N ₂ +CO ₂ +C ₃ H ₈ /CH ₄	N ₂ : 0.005 mol/mol to 0.02 mol/mol CO ₂ : 0.005 mol/mol to 0.02 mol/mol C ₃ H ₈ : 0.02 mol/mol to 0.1 mol/mol	N ₂ : 0.2 mmol/mol CO ₂ : 0.1 mmol/mol C ₃ H ₈ : 0.3 mmol/mol		N ₂ : •GC-TCD CO ₂ : •GC-TCD C ₃ H ₈ : •GC-TCD •GC-FID
		synthetic natural gas	N ₂ : 5 mmol/mol to 200 mmol/mol CO ₂ : 5 mmol/mol to 100 mmol/mol C ₂ H ₆ : 2 mmol/mol to 200 mmol/mol C ₃ H ₈ : 1 mmol/mol to 100 mmol/mol <i>n</i> -C ₄ H ₁₀ : 0.5 mmol/mol to 10 mmol/mol <i>iso</i> -C ₄ H ₁₀ : 0.5 mmol/mol to 10 mmol/mol CH ₄ : 600 mmol/mol to 980 mmol/mol	N ₂ : 0.5 % to 0.3 % (relative) CO ₂ : 0.6 % to 0.4 % (relative) C ₂ H ₆ : 0.5 % to 0.3 % (relative) C ₃ H ₈ : 0.5 % to 0.3 % (relative) <i>n</i> -C ₄ H ₁₀ : 0.5 % to 0.3 % (relative) <i>iso</i> -C ₄ H ₁₀ : 0.5 % to 0.3 % (relative) CH ₄ : 0.5 % to 0.3 % (relative)		N ₂ : •GC-TCD CO ₂ : •GC-TCD C ₂ H ₆ : •GC-FID •GC-TCD C ₃ H ₈ : •GC-FID •GC-TCD <i>n</i> -C ₄ H ₁₀ : •GC-FID •GC-TCD <i>iso</i> -C ₄ H ₁₀ : •GC-FID •GC-TCD CH ₄ : •GC-TCD •subtracting method
		N ₂ /Ar	1 μmol/mol to 200 μmol/mol	10 % to 0.5 % (relative)		• GC-MS
	CO ₂ /air	150 μmol/mol to 800 μmol/mol	0.02 μmol/mol to 0.1 μmol/mol	• CRDS		

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Inorganic standard solution	Mg	0.8 g/kg to 1.2 g/kg	0.16 % (relative)	• Chelatometric titration	2024-11-01
	Al	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Cu	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Zn	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Fe	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Ni	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Sr	0.8 g/kg to 1.2 g/kg	0.08 % (relative)	• Gravimetric preparation	
	V	0.8 g/kg to 1.2 g/kg	0.08 % (relative)	• Gravimetric preparation	
	Mn	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Mo	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Co	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Cd	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Ga	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	In	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Pb	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Bi	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Ba	0.8 g/kg to 1.2 g/kg	0.16 % (relative)	• Gravimetric preparation	
	Cr	0.8 g/kg to 1.2 g/kg	0.06 % (relative)	• Gravimetric preparation	
	Tl	0.8 g/kg to 1.2 g/kg	0.28 % (relative)	• Gravimetric preparation	
	Sn	0.8 g/kg to 1.2 g/kg	0.14 % (relative)	• Gravimetric preparation	
	Na	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	K	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Li	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Rb	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Cs	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	As	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Sb	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Be	0.8 g/kg to 1.2 g/kg	0.18 % (relative)	• Gravimetric preparation	
	Zr	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Ag	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
	Ca	0.8 g/kg to 1.2 g/kg	0.10 % (relative)	• Gravimetric preparation	
	Hg	0.8 g/kg to 1.2 g/kg	0.10 % (relative)	• Gravimetric preparation	
	Se	0.8 g/kg to 1.2 g/kg	0.12 % (relative)	• Gravimetric preparation	
	B	0.8 g/kg to 1.2 g/kg	0.12 % (relative)	• Gravimetric preparation	
	Te	0.8 g/kg to 1.2 g/kg	0.13 % (relative)	• Gravimetric preparation	
	Si	0.8 g/kg to 1.2 g/kg	0.28 % (relative)	• Gravimetric preparation	
	La	0.8 g/kg to 1.2 g/kg	0.13 % (relative)	• Chelatometric titration	
	Ti	0.8 g/kg to 1.2 g/kg	0.19 % (relative)	• Gravimetric preparation	
	Y	0.8 g/kg to 1.2 g/kg	0.13 % (relative)	• Chelatometric titration	
	chloride ion	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation	
nitrite ion	0.8 g/kg to 1.2 g/kg	0.18 % (relative)	• Gravimetric preparation		
nitrate ion	0.8 g/kg to 1.2 g/kg	0.15 % (relative)	• Gravimetric preparation		
phosphate ion	0.8 g/kg to 1.2 g/kg	0.18 % (relative)	• Gravimetric preparation		
bromide ion	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation		
iodide ion	0.8 g/kg to 1.2 g/kg	0.04 % (relative)	• Gravimetric preparation		
sulfate ion	0.8 g/kg to 1.2 g/kg	0.12 % (relative)	• IC		
cyanide ion	0.8 g/kg to 1.2 g/kg	1.1 % (relative)	• Complexometric titration		

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Inorganic standard solution	chlorate ion	0.8 g/kg to 1.2 g/kg	0.15 %(relative)	• Gravimetric titration	2024-11-01
	bromate ion	1.6 g/kg to 2.4 g/kg	0.14 %(relative)	• Gravimetric titration	
	ammonium ion	0.8 g/kg to 1.2 g/kg	0.13 %(relative)	• Gravimetric preparation	
	total organic carbon	0.8 g/kg to 1.2 g/kg	0.16 %(relative)	• Gravimetric preparation	
Inorganic standard solution (Isotopic standard)	²⁰⁶ Pb/ ²⁰⁴ Pb (Isotopic ratio)	14 mol/mol to 22 mol/mol	0.025 %(relative)	• MC-ICP-MS	
	²⁰⁷ Pb/ ²⁰⁴ Pb (Isotopic ratio)	13 mol/mol to 17 mol/mol	0.023 %(relative)	• MC-ICP-MS	
	²⁰⁸ Pb/ ²⁰⁴ Pb (Isotopic ratio)	36 mol/mol to 40 mol/mol	0.023 %(relative)	• MC-ICP-MS	
	²⁰⁸ Pb/ ²⁰⁶ Pb (Isotopic ratio)	1.8 mol/mol to 2.2 mol/mol	0.0062 %(relative)	• MC-ICP-MS	
	²⁰⁷ Pb/ ²⁰⁶ Pb (Isotopic ratio)	0.8 mol/mol to 1.0 mol/mol	0.0042 %(relative)	• MC-ICP-MS	
	²⁰⁴ Pb (Isotopic abundance)	0.012 mol/mol to 0.015 mol/mol	0.029 %(relative)	• MC-ICP-MS	
	²⁰⁶ Pb (Isotopic abundance)	0.24 mol/mol to 0.28 mol/mol	0.0036 %(relative)	• MC-ICP-MS	
	²⁰⁷ Pb (Isotopic abundance)	0.20 mol/mol to 0.23 mol/mol	0.0047 %(relative)	• MC-ICP-MS	
	²⁰⁸ Pb (Isotopic abundance)	0.51 mol/mol to 0.53 mol/mol	0.0031 %(relative)	• MC-ICP-MS	
	Pb (Molar mass)	207.1 g/mol to 207.3 g/mol	0.000014 %(relative)	• MC-ICP-MS	
Inorganic standard solution (Isotopic standard)	⁵⁶ Fe/ ⁵⁴ Fe (Isotopic ratio)	11 mol/mol to 20 mol/mol	0.041 %(relative)	• MC-ICP-MS	
	⁵⁷ Fe/ ⁵⁴ Fe (Isotopic ratio)	0.25 mol/mol to 0.47 mol/mol	0.063 %(relative)	• MC-ICP-MS	
	⁵⁸ Fe/ ⁵⁴ Fe (Isotopic ratio)	0.034 mol/mol to 0.063 mol/mol	0.11 %(relative)	• MC-ICP-MS	
	⁵⁴ Fe (Isotopic abundance)	0.041 mol/mol to 0.076 mol/mol	0.038 %(relative)	• MC-ICP-MS	
	⁵⁶ Fe (Isotopic abundance)	0.064 mol/mol to 1.2 mol/mol	0.0037 %(relative)	• MC-ICP-MS	
	⁵⁷ Fe (Isotopic abundance)	0.015 mol/mol to 0.028 mol/mol	0.071 %(relative)	• MC-ICP-MS	
	⁵⁸ Fe (Isotopic abundance)	0.0020 mol/mol to 0.0037 mol/mol	0.11 %(relative)	• MC-ICP-MS	
	Fe (Molar mass)	55.29 g/mol to 56.4 g/mol	0.000068 %(relative)	• MC-ICP-MS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method* ¹	Date of Accreditation
pH standard solution	pH	1.18 to 10.51	0.003	• Harned cell method	2024-11-01
Electrolytic conductivity standard solution	Electrolytic conductivity	0.05 S/m to 15 S/m	0.15 % to 0.48 % (relative)	• Impedance measurement	
		0.005 S/m to 0.05 S/m	0.61 % (relative)	• Impedance measurement	
High purity inorganic material (Potassium hydrogen phthalate)	acid	99.9 % to 100.1 % (mass fraction as potassium hydrogen phthalate)	0.012 % to 0.015 %	• Coulometric titration	
High purity inorganic material (Potassium dichromate)	oxidant	99.9 % to 100.1 % (mass fraction as potassium dichromate)	0.010 % to 0.012 %	• Coulometric titration	
High purity inorganic material (Arsenic(III) trioxide)	reductant	99.9 % to 100.1 % (mass fraction as arsenic(III) trioxide)	0.014 % to 0.020 %	• Coulometric titration	
High purity inorganic material (Sodium carbonate)	base	99.9 % to 100.1 % (mass fraction as sodium carbonate)	0.01 % to 0.02 %	• Coulometric titration • Gravimetric titration	
High purity inorganic material (Potassium iodate)	oxidant	99.9 % to 100.1 % (mass fraction as potassium iodate)	0.014 % to 0.020 %	• Coulometric titration • Gravimetric titration	
High purity inorganic material (Sodium oxalate)	reductant	99.9 % to 100.1 % (mass fraction as sodium oxalate)	0.023 % to 0.025 %	• Coulometric titration • Gravimetric titration	
Heavy metals in polymer	Cd	5 mg/kg to 10000 mg/kg	0.5 % to 2.0 % (relative)	• ICP-OES • ICP-MS • ID-ICP-MS	
	Cr	10 mg/kg to 10000 mg/kg	0.5 % to 2.0 % (relative)	• ICP-OES • ICP-MS • ID-ICP-MS	
	Hg	10 mg/kg to 10000 mg/kg	0.5 % to 2.0 % (relative)	• ICP-OES • ICP-MS • ID-ICP-MS	
	Pb	10 mg/kg to 10000 mg/kg	0.5 % to 2.0 % (relative)	• ICP-OES • ICP-MS • ID-ICP-MS	
	Br	50 mg/kg to 10000 mg/kg	2.0 % to 5.0 % (relative)	• Instrumental Neutron Activation Analysis • ID-ICP-MS	
Minor elements in metals and alloys (lead-free solder)	Pb	100 mg/kg to 2000 mg/kg	0.8 % to 1.6 % (relative)	• ID-ICP-MS	
	Ag	2.8 % to 3.2 % (mass fraction)	0.8 % to 1.6 % (relative)	• ID-ICP-MS	
	Cu	0.3 % to 0.7 % (mass fraction)	0.5 % to 1.0 % (relative)	• ID-ICP-MS	
High purity inorganic material (Sodium chloride)	Cl	99.9 % to 100.1 % (mass fraction as sodium chloride)	0.03 % to 0.05 %	• Coulometric titration	
High purity inorganic material (Ammonium chloride)	ammonium ion	99.9 % to 100.1 % (mass fraction as ammonium chloride)	0.034 % to 0.070 %	• Coulometric titration	
	Cl	99.9 % to 100.1 % (mass fraction as ammonium chloride)	0.054 % to 0.080 %	• Gravimetric titration	
High purity inorganic material (Amidosulfuric acid)	acid	99.9 % to 100.1 % (mass fraction as amidosulfuric acid)	0.008 % to 0.012 %	• Coulometric titration	
	N	99.9 % to 100.1 % (mass fraction as amidosulfuric acid)	0.025 % to 0.040 %	• Coulometric titration	
Hydrochloric acid	acid	0.05 mol/kg to 2 mol/kg	0.016 % to 0.027 % (relative)	• Coulometric titration	
High purity inorganic material (Tris(hydroxymethyl)aminomethane)	base	99.8 % to 100.2 % (mass fraction as tris(hydroxymethyl)aminomethane)	0.026 %	• Coulometric titration	
High purity inorganic material (Calcium carbonate)	Ca	99.5 % to 100.5 % (mass fraction as calcium carbonate)	0.030 %	• Chelatometric titration	
High purity inorganic material (Zinc)	Zn	99.5 % to 100.0 % (mass fraction as zinc)	0.008 %	• Subtracting method with impurity analysis	
	Zn (molar mass)	65.36 g/mol to 65.40 g/mol	0.0018 % (relative)	• ICP-MS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
High purity organic materials	ethanol	0.998 mol/mol to 1 mol/mol	0.002 mol/mol to 0.0004 mol/mol	• Freezing point depression method	2024-11-01
	toluene	0.998 mol/mol to 1 mol/mol	0.003 mol/mol to 0.00006 mol/mol	• Freezing point depression method	
	1,2-dichloroethane	0.998 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0001 mol/mol	• Freezing point depression method	
	benzene	0.998 mol/mol to 1 mol/mol	0.001 mol/mol to 0.00002 mol/mol	• Freezing point depression method	
	<i>o</i> -xylene	0.998 mol/mol to 1 mol/mol	0.001 mol/mol to 0.00002 mol/mol	• Freezing point depression method	
	ethylbenzene	0.998 mol/mol to 1 mol/mol	0.0002 mol/mol to 0.002 mol/mol	• Freezing point depression method	
	cholesterol	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Freezing point depression method	
	<i>m</i> -xylene	0.997 mol/mol to 1 mol/mol	0.001 mol/mol to 0.00015 mol/mol	• Freezing point depression method	
	diethyl phthalate	0.997 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0002 mol/mol	• Freezing point depression method	
	chloroform	0.995 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0002 mol/mol	• Freezing point depression method	
	<i>p</i> -xylene	0.995 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0001 mol/mol	• Freezing point depression method	
	bromoform	0.995 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0002 mol/mol	• Freezing point depression method	
	bromodichloromethane	0.995 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0002 mol/mol	• Freezing point depression method	
	bisphenol A	0.995 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0006 mol/mol	• Freezing point depression method	
	dibromochloromethane	0.995 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0002 mol/mol	• Freezing point depression method	
	<i>trans</i> -1,2-dichloroethylene	0.995 mol/mol to 1 mol/mol	0.001 mol/mol to 0.0002 mol/mol	• Freezing point depression method	
	trichloroethylene	0.995 mol/mol to 1 mol/mol	0.002 mol/mol	• Freezing point depression method	
	tetrachloroethylene	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0001 mol/mol	• Freezing point depression method	
	1,1,1-trichloroethane	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0004 mol/mol	• Freezing point depression method	
	<i>cis</i> -1,2-dichloroethylene	0.99 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0007 mol/mol	• Freezing point depression method	
	<i>cis</i> -1,3-dichloropropene	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.003 mol/mol	• Freezing point depression method	
	1,4-dichlorobenzene	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0003 mol/mol	• Freezing point depression method	
styrene	0.99 kg/kg to 1.00 kg/kg	0.01 kg/kg to 0.0005 kg/kg	• Freezing point depression method • Subtracting method		

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
High purity organic materials	dichloromethane	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0001 mol/mol	• Freezing point depression method	2024-11-01
	tetrachloromethane	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0001 mol/mol	• Freezing point depression method	
	1,1-dichloroethylene	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0001 mol/mol	• Freezing point depression method	
	1,1,2-trichloroethane	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.0001 mol/mol	• Freezing point depression method	
	<i>trans</i> -1,3-dichloropropene	0.97 mol/mol to 1 mol/mol	0.005 mol/mol to 0.003 mol/mol	• Freezing point depression method	
	1,2-dichloropropane	0.995 mol/mol to 1 mol/mol	0.005 mol/mol to 0.003 mol/mol	• Freezing point depression method	
	acrylonitrile	0.99 kg/kg to 1.00 kg/kg	0.01 kg/kg to 0.00005 kg/kg	• Freezing point depression method • Subtracting method	
	acetaldehyde	0.99 kg/kg to 1.00 kg/kg	0.01 kg/kg to 0.003 kg/kg	• Titration • Subtracting method	
	17 β -estradiol	0.96 kg/kg to 1.00 kg/kg	0.005 kg/kg to 0.003 kg/kg	• qNMR • Subtracting method (HPLC-UV, HPLC-CAD, HS-GC-MS, Coulometric Karl-Fischer titration, TG)	
	progesterone	0.98 kg/kg to 1.00 kg/kg	0.01 kg/kg to 0.001 kg/kg	• qNMR • Freezing point depression method • Subtracting method (HPLC-UV, HPLC-CAD, HS-GC-MS, Coulometric Karl-Fischer titration, TG)	
	testosterone	0.98 kg/kg to 1.00 kg/kg	0.01 kg/kg to 0.001 kg/kg	• qNMR • Subtracting method (HPLC-UV, HPLC-CAD, HS-GC-MS, Coulometric Karl-Fischer titration, TG)	
	sulfur in organic materials (as sulfur)	0.2 kg/kg to 0.4 kg/kg	0.00006 kg/kg to 0.0004 kg/kg	• Freezing point depression method • Subtracting method (GC-FID, GC-SCD, Coulometric Karl-Fischer titration)	
	dibutyl sulfide	0.995 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0001 kg/kg	• Freezing point depression method • Subtracting method (GC-FID, GC-SCD, Coulometric Karl-Fischer titration)	
	1,4-dioxane	0.998 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0001 kg/kg	• Freezing point depression method	
	<i>tert</i> -butylmethylether	0.998 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0003 kg/kg	• Freezing point depression method	
	trichloroacetic acid	0.995 kg/kg to 1 kg/kg	0.002 kg/kg	• Freezing point depression method • Titration	
	3,5-bis(trifluoromethyl)benzoic acid	0.999 kg/kg to 1 kg/kg	0.0003 kg/kg to 0.0001 kg/kg	• Freezing point depression method • Coulometric titration • Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration, TG)	
1,4-bis(trimethylsilyl)-2,3,5,6-tetrafluorobenzene	0.999 kg/kg to 1 kg/kg	0.0003 kg/kg to 0.0001 kg/kg	• Freezing point depression method • Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration, TG)		

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
High purity organic materials	di- <i>n</i> -butyl phthalate	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	2024-11-01
	di-2-ethylhexyl phthalate	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	di- <i>n</i> -propyl phthalate	0.98 kg/kg to 1 kg/kg	0.0006 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	di- <i>n</i> -pentyl phthalate	0.97 kg/kg to 1 kg/kg	0.006 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	di- <i>n</i> -hexyl phthalate	0.97 kg/kg to 1 kg/kg	0.006 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	dicyclohexyl phthalate	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	butyl benzyl phthalate	0.98 kg/kg to 1 kg/kg	0.0015 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	simazine	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, GC-MS, Coulometric Karl-Fischer titration)	
	thiuram	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• qNMR • Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	thiobencarb	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• Freezing point depression method • qNMR • Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	4- <i>n</i> -nonylphenol	0.99 mol/mol to 1 mol/mol	0.005 mol/mol to 0.001 mol/mol	• Freezing point depression method	
	4- <i>t</i> -octylphenol	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	4- <i>t</i> -butylphenol	0.98 kg/kg to 1 kg/kg	0.001 kg/kg to 0.0002 kg/kg	• Subtracting method (HPLC-UV, GC-FID, Coulometric Karl-Fischer titration)	
	4- <i>n</i> -heptylphenol	0.99 mol/mol to 1 mol/mol	0.005 mol/mol to 0.001 mol/mol	• Freezing point depression method	
	2,4-dichlorophenol	0.99 mol/mol to 1 mol/mol	0.005 mol/mol to 0.001 mol/mol	• Freezing point depression method	
Environmental matrix (fish oil)	<i>p,p'</i> -DDE	1 mg/kg to 10 mg/kg	0.014 mg/kg	• ID-GC-MS	
	<i>p,p'</i> -DDT	0.05 mg/kg to 0.5 mg/kg	0.0031 mg/kg	• ID-GC-MS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Organic standard solution	<i>p,p'</i> -DDT/2,2,4-trimethylpentane	0.05 mg/kg to 20 mg/kg	7 % (relative)	• Freezing point depression method • HPLC-UV • Gravimetric preparation	2024-11-01
	<i>p,p'</i> -DDE/2,2,4-trimethylpentane	0.5 mg/kg to 20 mg/kg	2 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	γ -HCH/2,2,4-trimethylpentane	0.03 mg/kg to 20 mg/kg	1 % (relative)	• Subtracting method (GC-FID) • Gravimetric preparation	
	<i>p,p'</i> -DDT + <i>p,p'</i> -DDE + <i>p,p'</i> -DDD + γ -HCH /2,2,4-trimethylpentane	<i>p,p'</i> -DDT : 0.05 mg/kg to 20 mg/kg <i>p,p'</i> -DDE : 0.5 mg/kg to 20 mg/kg <i>p,p'</i> -DDD : 0.5 mg/kg to 20 mg/kg γ -HCH : 0.03 mg/kg to 20 mg/kg	<i>p,p'</i> -DDT : 2 % to 1 % (relative) <i>p,p'</i> -DDE : 1 % to 0.5 % (relative) <i>p,p'</i> -DDD : 1 % to 0.5 % (relative) γ -HCH : 2 % to 0.5 % (relative)	• Freezing point depression method • HPLC-UV • GC-FID • Gravimetric preparation	
	PCB28/2,2,4-trimethylpentane	2 mg/kg to 50 mg/kg	1.7 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	PCB70/2,2,4-trimethylpentane	2 mg/kg to 50 mg/kg	1.8 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	PCB105/2,2,4-trimethylpentane	2 mg/kg to 50 mg/kg	2.4 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	PCB153/2,2,4-trimethylpentane	2 mg/kg to 50 mg/kg	1.7 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	PCB170/2,2,4-trimethylpentane	2 mg/kg to 50 mg/kg	2.0 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	PCB194/2,2,4-trimethylpentane	2 mg/kg to 50 mg/kg	1.6 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	PCB28+PCB70+PCB105+PCB153+PCB170+PCB194 /2,2,4-trimethylpentane	PCB28 : 2 mg/kg to 50 mg/kg PCB70 : 2 mg/kg to 50 mg/kg PCB105 : 2 mg/kg to 50 mg/kg PCB153 : 2 mg/kg to 50 mg/kg PCB170 : 2 mg/kg to 50 mg/kg PCB194 : 2 mg/kg to 50 mg/kg	PCB28 : 1.7 % (relative) PCB70 : 1.8 % (relative) PCB105 : 2.4 % (relative) PCB153 : 1.7 % (relative) PCB170 : 2.0 % (relative) PCB194 : 1.6 % (relative)	• Freezing point depression method • GC-FID • Gravimetric preparation	
	4-hydroxy-clomifene	4-hydroxy-clomifene: 200 μ g/g to 300 μ g/g (<i>E</i>)-4-hydroxy-clomifene: 50 μ g/g to 200 μ g/g (<i>Z</i>)-4-hydroxy-clomifene: 50 μ g/g to 200 μ g/g	4-hydroxy-clomifene: 1.5 % (relative) (<i>E</i>)-4-hydroxy-clomifene: 1.6 % (relative) (<i>Z</i>)-4-hydroxy-clomifene: 1.6 % (relative)	• qNMR • qNMR/HPLC-UV • Gravimetric preparation	
	3 β ,4 α -dihydroxy-5 α -androstan-17-one	100 μ g/g to 170 μ g/g	1.4 % (relative)	• qNMR • qNMR/HPLC-UV • Gravimetric preparation	
	sulfur in toluene (as sulfur)	0.5 mg/kg to 10000 mg/kg	0.02 mg/kg to 10 mg/kg	• Freezing point depression method • Subtracting method (GC-FID, GC-FPD, Coulometric Karl-Fischer titration) • Gravimetric preparation	
		10 μ g/kg to 500 μ g/kg	5 μ g/kg to 20 μ g/kg	• Combustion-ultraviolet fluorescence method	
CRMs for thermal properties	cyclohexane (thermal analysis with thermal analyzer such as DSC)	phase transition temperature 186 K to 280 K	0.04 K to 0.1 K	• Adiabatic calorimetry	
		phase transition enthalpy 30 J g ⁻¹ to 90 J g ⁻¹	0.7 J g ⁻¹ to 3 J g ⁻¹	• Adiabatic calorimetry	
High purity organic materials	perfluorooctanoic acid	0.95 kg/kg to 1 kg/kg	0.006 kg/kg to 0.002 kg/kg	• Titration • Subtracting method (LC-MS, Karl Fischer titration, TG)	
	chloroalkanes	0.98 kg/kg to 1 kg/kg	0.005 kg/kg to 0.001 kg/kg	• Subtracting method (GC-FID, HS-GC-MS, Karl Fischer titration, TG)	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Organic standard solution	benzo[a]pyrene/ 2,2,4-trimethylpentane	10 mg/kg to 200 mg/kg	4 % to 1 % (relative)	• Freezing point depression method • Gravimetric preparation	2024-11-01
	potassium perfluorooctanesulfonate /methanol	5 mg/kg to 100 mg/kg	4 % to 1 % (relative)	• Freezing point depression method • Gravimetric preparation	
Standard solution (water in organic solvent)	water	0.01 g/kg to 10 g/kg	30 % to 0.1 % (relative)	• Coulometric titration • Volumetric titration	
Food (pesticide in grain)	fenitrothion	0.1 mg/kg to 1 mg/kg	20 % to 5 % (relative)	• ID-GC-MS • ID-LC-MS	
	etofenprox	0.1 mg/kg to 1 mg/kg	30 % to 5 % (relative)	• ID-GC-MS • ID-LC-MS	
Food (pesticide in vegetable)	diazinon	0.1 mg/kg to 100 mg/kg	40 % to 5 % (relative)	• ID-GC-MS	
	fenitrothion	0.1 mg/kg to 100 mg/kg	20 % to 3 % (relative)	• ID-GC-MS	
	chlorpyrifos	1 mg/kg to 100 mg/kg	40 % to 5 % (relative)	• ID-GC-MS	
	permethrin	0.1 mg/kg to 100 mg/kg	30 % to 4 % (relative)	• ID-GC-MS	
	cypermethrin	0.1 mg/kg to 100 mg/kg	40 % to 5 % (relative)	• ID-GC-MS	
	etofenprox	1 mg/kg to 100 mg/kg	20 % to 3 % (relative)	• ID-GC-MS	
Food (pesticide in fruits)	diazinon	0.1 mg/kg to 10 mg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	fenitrothion	0.1 mg/kg to 10 mg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	permethrin	0.1 mg/kg to 10 mg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	cypermethrin	0.1 mg/kg to 10 mg/kg	30 % to 3 % (relative)	• ID-GC-MS	
Food (pesticide in beans)	diazinon	0.001 mg/kg to 0.1 mg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	fenitrothion	0.001 mg/kg to 0.2 mg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	chlorpyrifos	0.001 mg/kg to 0.3 mg/kg	30 % to 3 % (relative)	• ID-GC-MS	
	permethrin	0.002 mg/kg to 0.1 mg/kg	20 % to 2 % (relative)	• ID-GC-MS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Environmental matrix (trace elements in sediment)	Sb	0.1 mg/kg to 3 mg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	2024-11-01
	Cd	0.1 mg/kg to 3 mg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • GFAAS 	
	Cu	5 mg/kg to 500 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-OES • GFAAS 	
	Pb	2 mg/kg to 250 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-OES • GFAAS 	
	Ni	5 mg/kg to 50 mg/kg	5 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-OES • GFAAS 	
	Zn	20 mg/kg to 1000 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-OES 	
	As	1 mg/kg to 50 mg/kg	20 % to 2 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-OES • GFAAS • HR-ICP-MS 	
	Co	1 mg/kg to 50 mg/kg	15 % to 2 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-OES • GFAAS 	
	Se	0.1 mg/kg to 5 mg/kg	20 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • HR-ICP-MS 	
	Cr	10 mg/kg to 500 mg/kg	10 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-OES • GFAAS 	
	Hg	0.02 mg/kg to 5 mg/kg	15 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • Heating evaporation-Gold amalgamation AAS 	
	Ag	0.05 mg/kg to 2 mg/kg	4 % to 3 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Mo	0.5 mg/kg to 20 mg/kg	7 % to 3 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Sn	1 mg/kg to 50 mg/kg	5 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
Environmental (polychlorinated biphenyls in mineral oil)	PCB3	0.2 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB8	0.2 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB28	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB52	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB101	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB118	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB138	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB153	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB180	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB194	0.1 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	
	PCB206	0.09 µg/kg to 10 mg/kg	50 % to 3 % (relative)	• ID-GC-MS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method* ¹	Date of Accreditation
Environmental matrix (fish tissue)	PCB28	1 µg/kg to 100 µg/kg	15 % to 2 % (relative)	• ID-GC-MS	2024-11-01
	PCB70	1 µg/kg to 10 µg/kg	15 % to 5 % (relative)	• ID-GC-MS	
	PCB105	1 µg/kg to 100 µg/kg	15 % to 2 % (relative)	• ID-GC-MS	
	PCB153	10 µg/kg to 200 µg/kg	10 % to 2 % (relative)	• ID-GC-MS	
	PCB170	0.1 µg/kg to 10 µg/kg	10 % to 4 % (relative)	• ID-GC-MS	
	<i>p,p'</i> -DDT	1 µg/kg to 10 µg/kg	10 % to 5 % (relative)	• ID-GC-MS	
	<i>p,p'</i> -DDE	10 µg/kg to 100 µg/kg	15 % to 5 % (relative)	• ID-GC-MS	
	<i>p,p'</i> -DDD	1 µg/kg to 10 µg/kg	10 % to 5 % (relative)	• ID-GC-MS	
	dieldrin	1 µg/kg to 10 µg/kg	10 % to 3 % (relative)	• ID-GC-MS	
	<i>trans</i> -nonachlor	1 µg/kg to 10 µg/kg	10 % to 4 % (relative)	• ID-GC-MS	
Environmental matrix (PAHs/dust)	fluorene	0.1 mg/kg to 100 mg/kg	40 % to 10 % (relative)	• ID-GC-MS	
	anthracene	0.1 mg/kg to 100 mg/kg	40 % to 10 % (relative)	• ID-GC-MS	
	fluoranthene	1 mg/kg to 1000 mg/kg	30 % to 10 % (relative)	• ID-GC-MS	
	pyrene	1 mg/kg to 1000 mg/kg	30 % to 10 % (relative)	• ID-GC-MS	
	benzo[<i>a</i>]anthracene	0.1 mg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-GC-MS	
	benzo[<i>b</i>]fluoranthene	0.1 mg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-GC-MS	
	benzo[<i>k</i>]fluoranthene	0.01 mg/kg to 10 mg/kg	20 % to 10 % (relative)	• ID-GC-MS	
	benzo[<i>a</i>]pyrene	0.1 mg/kg to 100 mg/kg	30 % to 10 % (relative)	• ID-GC-MS	
	perylene	0.01 mg/kg to 10 mg/kg	30 % to 10 % (relative)	• ID-GC-MS	
	indeno[1,2,3- <i>cd</i>]pyrene	0.1 mg/kg to 100 mg/kg	40 % to 10 % (relative)	• ID-GC-MS	
benzo[<i>ghi</i>]perylene	0.1 mg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-GC-MS		
Environmental matrix (toxic elements in tunnel dust)	Cr	5 mg/kg to 5 % (mass fraction)	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS	
	Ni	5 mg/kg to 2 % (mass fraction)	5 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES	
	Pb	2 mg/kg to 1 % (mass fraction)	5 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES	
	Mn	2 mg/kg to 1 % (mass fraction)	5 % to 2 % (relative)	• ICP-MS • ICP-OES • GFAAS	
	Cd	0.1 mg/kg to 0.1 % (mass fraction)	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS	
Environmental matrix (polychlorinated biphenyls / pesticide in biological sample)	PCB118	5 ng/kg to 200 ng/kg	40 % to 10 % (relative)	• ID-GC-MS	
	PCB138	5 ng/kg to 200 ng/kg	40 % to 10 % (relative)	• ID-GC-MS	
	PCB153	5 ng/kg to 200 ng/kg	40 % to 10 % (relative)	• ID-GC-MS	
	PCB194	5 ng/kg to 200 ng/kg	40 % to 10 % (relative)	• ID-GC-MS	
	acetamiprid	0.1 µg/kg to 2 µg/kg	50 % to 10 % (relative)	• ID-LC-MS	
	clothianidin	0.1 µg/kg to 2 µg/kg	50 % to 10 % (relative)	• ID-LC-MS	
	thiacloprid	0.1 µg/kg to 2 µg/kg	50 % to 10 % (relative)	• ID-LC-MS	
	thiamethoxam	0.1 µg/kg to 2 µg/kg	50 % to 10 % (relative)	• ID-LC-MS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Environmental (polychlorinated biphenyls and organochlorine pesticides in sediment)	PCB3	0.1 µg/kg to 100 µg/kg	30 % to 5 % (relative)	• ID-GC-MS	2024-11-01
	PCB15	0.1 µg/kg to 100 µg/kg	20 % to 4 % (relative)	• ID-GC-MS	
	PCB28	1 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB31	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB70	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB101	1 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB105	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB138	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB153	1 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB170	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB180	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB194	0.1 µg/kg to 100 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB206	0.1 µg/kg to 100 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	PCB209	0.1 µg/kg to 100 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	<i>p,p'</i> -DDT	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	<i>p,p'</i> -DDE	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	<i>p,p'</i> -DDD	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
	γ -HCH	0.5 µg/kg to 1000 µg/kg	20 % to 2 % (relative)	• ID-GC-MS	
Environmental (polycyclic aromatic hydrocarbons in sediment)	fluorene	1 µg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-GC-MS	2024-11-01
	phenanthrene	1 µg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	anthracene	1 µg/kg to 100 mg/kg	40 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	fluoranthene	1 µg/kg to 100 mg/kg	20 % to 5 % (relative)	• ID-GC-MS • ID-LC-MS	
	pyrene	1 µg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	benzo[<i>c</i>]phenanthrene	1 µg/kg to 100 mg/kg	10 % to 5 % (relative)	• ID-GC-MS • ID-LC-MS	
	benz[<i>a</i>]anthracene	1 µg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	chrysene	1 µg/kg to 100 mg/kg	10 % to 5 % (relative)	• ID-GC-MS • ID-LC-MS	
	benzo[<i>b</i>]fluoranthene	1 µg/kg to 100 mg/kg	40 % to 10 % (relative)	• ID-GC-MS	
	benzo[<i>j</i>]fluoranthene	1 µg/kg to 100 mg/kg	40 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	benzo[<i>k</i>]fluoranthene	1 µg/kg to 100 mg/kg	30 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	benzo[<i>a</i>]fluoranthene	1 µg/kg to 100 mg/kg	50 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	benzo[<i>e</i>]pyrene	1 µg/kg to 100 mg/kg	30 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	benzo[<i>a</i>]pyrene	1 µg/kg to 100 mg/kg	20 % to 5 % (relative)	• ID-GC-MS • ID-LC-MS	
	perylene	100 µg/kg to 100 mg/kg	30 % to 10 % (relative)	• ID-GC-MS	
	indeno[1,2,3- <i>cd</i>]pyrene	1 µg/kg to 100 mg/kg	40 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	benzo[<i>ghi</i>]perylene	1 µg/kg to 100 mg/kg	30 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
	dibenz[<i>a,h</i>]anthracene	1 µg/kg to 100 mg/kg	50 % to 10 % (relative)	• ID-GC-MS • ID-LC-MS	
Fuel (components in bioethanol fuel)	water	100 mg/kg to 5000 mg/kg	2 % to 0.2 % (relative)	• Coulometric titration • Volumetric titration	2024-11-01
	methanol	0.2 g/kg to 1 g/kg	10 % to 2 % (relative)	• ID-GC-MS • GC-FID	
	S	1 mg/kg to 5 mg/kg	3 % (relative)	• Combustion-ultraviolet fluorescence method • Combustion-IC	
	Cu	0.0001 mg/kg to 500 mg/kg	10 % to 1 % (relative)	• ICP-MS • ID-ICP-MS • GFAAS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method* ¹	Date of Accreditation
Fuel (components in biodiesel fuel)	water	300 mg/kg to 1000 mg/kg	10 % to 5 % (relative)	<ul style="list-style-type: none"> • Coulometric titration • Volumetric titration 	2024-11-01
	Na	0.5 mg/kg to 20 mg/kg	20 % to 5 % (relative)	<ul style="list-style-type: none"> • ICP-MS/MS • HR-ICP-MS • FAAS 	
	Mg	0.5 mg/kg to 20 mg/kg	20 % to 5 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS/MS • ICP-MS/MS 	
	K	0.5 mg/kg to 20 mg/kg	20 % to 5 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS/MS • ICP-MS/MS 	
	Ca	0.5 mg/kg to 20 mg/kg	20 % to 5 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS/MS • ICP-MS/MS 	
	P	0.5 mg/kg to 20 mg/kg	20 % to 5 % (relative)	<ul style="list-style-type: none"> • ICP-MS/MS • FI-ICP-MS • ICP-OES 	
	S	2 mg/kg to 50 mg/kg	10 % to 5 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS/MS • ICP-MS/MS • Combustion-IC 	
Environmental matrix (river water and drinking water)	Al	1 µg/kg to 100 µg/kg	8 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-MS/MS • GFAAS 	2024-11-01
	Sb	0.001 µg/kg to 10 µg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • HR-ICP-MS • ICP-MS/MS 	
	As	0.05 µg/kg to 50 µg/kg	15 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-MS/MS • GFAAS 	
	Ba	0.5 µg/kg to 50 µg/kg	2 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	B	1 µg/kg to 100 µg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Cd	0.001 µg/kg to 10 µg/kg	15 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Cr	0.05 µg/kg to 50 µg/kg	8 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Cu	0.05 µg/kg to 50 µg/kg	15 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Fe	0.1 µg/kg to 100 µg/kg	10 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Pb	0.001 µg/kg to 10 µg/kg	15 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Mn	0.01 µg/kg to 50 µg/kg	15 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-MS/MS • GFAAS 	
	Mo	0.05 µg/kg to 10 µg/kg	2 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • HR-ICP-MS • ICP-MS/MS 	
	Ni	0.01 µg/kg to 50 µg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Se	0.1 µg/kg to 50 µg/kg	10 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Zn	0.05 µg/kg to 50 µg/kg	10 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	Na	1 mg/kg to 50 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-OES • MP-AES 	
	K	0.2 mg/kg to 50 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-OES • MP-AES 	
	Mg	0.2 mg/kg to 50 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-OES • MP-AES 	
	Ca	1 mg/kg to 50 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS • ICP-OES • MP-AES 	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method* ¹	Date of Accreditation
Environmental matrix (river water and drinking water)	Rb	0.05 µg/kg to 100 µg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	2024-11-01
	Sr	0.05 µg/kg to 200 µg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS • ICP-MS/MS 	
	P	1 µg/kg to 100 µg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • ICP-MS 	
Environmental matrix (sea water)	Cr	1 µg/kg to 20000 µg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Mn	1 µg/kg to 20000 µg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ICP-MS • GFAAS 	
	Fe	1 µg/kg to 20000 µg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Ni	1 µg/kg to 20000 µg/kg	15 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Cu	1 µg/kg to 20000 µg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Zn	1 µg/kg to 20000 µg/kg	20 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	As	1 µg/kg to 20000 µg/kg	15 % to 2 % (relative)	<ul style="list-style-type: none"> • ICP-MS • GFAAS 	
	Se	1 µg/kg to 20000 µg/kg	15 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Cd	0.3 µg/kg to 20000 µg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	Pb	1 µg/kg to 20000 µg/kg	10 % to 2 % (relative)	<ul style="list-style-type: none"> • ID-ICP-MS • ICP-MS 	
	dissolved silica	0.03 mg/kg to 5 mg/kg	12 % to 1 % (relative)	<ul style="list-style-type: none"> • Colorimetry • IC • IC-ID-ICP-MS 	
	nitrate ion	0.8 mg/kg to 3 mg/kg	3 % to 1 % (relative)	<ul style="list-style-type: none"> • Colorimetry • IC 	
nitrite ion	0.01 mg/kg to 0.3 mg/kg	20 % to 5 % (relative)	<ul style="list-style-type: none"> • Colorimetry • IC 		
phosphate ion	0.1 mg/kg to 0.3 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • Colorimetry 		
Standard solution for chemical speciation	arsenobetaine	1 mg/kg to 1000 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • HPLC-ICP-MS • ICP-MS • ICP-OES • GFAAS 	
	arsenate (As(V))	1 mg/kg to 1000 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • HPLC-ICP-MS • ICP-MS • ICP-OES • GFAAS 	
	dimethylarsenic acid	1 mg/kg to 1000 mg/kg	5 % to 1 % (relative)	<ul style="list-style-type: none"> • HPLC-ICP-MS • ICP-MS • ICP-OES • GFAAS 	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Food (trace elements and arsenic compounds in grains and beans)	Cr	0.01 mg/kg to 10 mg/kg	15 % to 2 % (relative)	• ID-HR-ICP-MS • ICP-MS	2024-11-01
	Mn	0.1 mg/kg to 50 mg/kg	10 % to 1.5 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • GFAAS • MP-AES	
	Fe	0.1 mg/kg to 100 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	Ni	0.01 mg/kg to 10 mg/kg	15 % to 2 % (relative)	• ID-ICP-MS • ICP-MS	
	Cu	0.1 mg/kg to 50 mg/kg	10 % to 1.5 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	Zn	0.1 mg/kg to 100 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	As	0.005 mg/kg to 50 mg/kg	10 % to 2 % (relative)	• ICP-MS • HR-ICP-MS • GFAAS	
	Rb	0.1 mg/kg to 50 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS	
	Sr	0.02 mg/kg to 10 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS	
	Cd	0.005 mg/kg to 5 mg/kg	7 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	Mo	0.02 mg/kg to 10 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS	
	Ba	0.02 mg/kg to 10 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS	
	Pb	0.001 mg/kg to 10 mg/kg	15 % to 2 % (relative)	• ID-HR-ICP-MS • ICP-MS	
	Na	0.1 mg/kg to 50 mg/kg	15 % to 2 % (relative)	• ICP-OES • FAAS • Flame photometry	
	Mg	10 mg/kg to 5000 mg/kg	5 % to 1.2 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS • MP-AES	
	K	100 mg/kg to 50000 mg/kg	5 % to 2 % (relative)	• ICP-OES • FAAS • Flame photometry	
	Ca	5 mg/kg to 5000 mg/kg	5 % to 1.5 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS • Flame photometry • MP-AES	
	P	100 mg/kg to 9000 mg/kg	10 % to 2 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES	
	arsenite (As(III))	0.005 mg/kg to 50 mg/kg (as As)	8 % to 2 % (relative)	• HPLC-ICP-MS	
	arsenate (As(V))	0.005 mg/kg to 50 mg/kg (as As)	8 % to 2 % (relative)	• HPLC-ICP-MS	
dimethylarsenic acid	0.005 mg/kg to 50 mg/kg (as As)	8 % to 2 % (relative)	• HPLC-ICP-MS		
Food (trace elements, arsenobetaine and methylmercury in fish, shellfish, and cephalopoda tissues)	Cr	0.2 mg/kg to 5 mg/kg	15 % to 3 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • GFAAS	
	Mn	0.1 mg/kg to 5 mg/kg	10 % to 1.5 % (relative)	• ICP-MS • HR-ICP-MS • GFAAS	
	Fe	1 mg/kg to 100 mg/kg	10 % to 3 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	Ni	0.2 mg/kg to 20 mg/kg	15 % to 3 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • GFAAS	
	Cu	0.2 mg/kg to 100 mg/kg	10 % to 1.5 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	Zn	1 mg/kg to 100 mg/kg	10 % to 1.5 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES	
	As	1 mg/kg to 100 mg/kg	10 % to 2 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • GFAAS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Food (trace elements, arsenobetaine and methylmercury in fish, shellfish, and cephalopoda tissues)	Se	0.1 mg/kg to 10 mg/kg	15 % to 3 % (relative)	• ID-ICP-MS • ICP-MS • GFAAS	2024-11-01
	Hg	0.1 mg/kg to 10 mg/kg	10 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • Heating evaporation Gold amalgamation AAS	
	Na	1 mg/kg to 100 g/kg	10 % to 2 % (relative)	• ICP-OES • FAAS • Flame photometry	
	Mg	0.5 mg/kg to 100 g/kg	5 % to 1 % (relative)	• ICP-MS • ICP-OES • FAAS	
	K	1 mg/kg to 100 g/kg	10 % to 2 % (relative)	• ICP-OES • FAAS • Flame photometry	
	Ca	0.1 mg/kg to 100 g/kg	15 % to 3 % (relative)	• ICP-MS • ICP-OES • FAAS • Flame photometry	
	arsenobetaine	1 mg/kg to 100 mg/kg (as As)	10 % to 2 % (relative)	• HPLC-ICP-MS • ID-LC-MS	
	methylmercury	0.1 mg/kg to 10 mg/kg (as Hg)	5 % to 1 % (relative)	• ID-GC-ICP-MS	
	Sr	0.02 mg/kg to 10 mg/kg	10 % to 1.2 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	Cd	0.01 mg/kg to 5 mg/kg	10 % to 1.5 % (relative)	• ID-ICP-MS • ID-HR-ICP-MS • ICP-MS • ICP-OES • GFAAS	
	P	1 g/kg to 100 g/kg	5 % to 2 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES	
	Food (trace elements and arsenic compounds in algae)	Na	0.5 g/kg to 100 g/kg	10 % to 1 % (relative)	
K		1 g/kg to 100 g/kg	10 % to 1 % (relative)	• ICP-OES • FAAS • Flame photometry	
Mg		0.1 g/kg to 100 g/kg	10 % to 1 % (relative)	• ICP-MS • ICP-OES • FAAS	
Ca		0.5 g/kg to 100 g/kg	10 % to 1 % (relative)	• ICP-MS • ICP-OES • FAAS • Flame photometry	
Sr		0.1 g/kg to 50 g/kg	10 % to 1 % (relative)	• ICP-MS • ID-ICP-MS • ICP-OES • GFAAS	
P		0.01 g/kg to 50 g/kg	10 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES	
Al		10 mg/kg to 1000 mg/kg	10 % to 3 % (relative)	• ICP-MS • ICP-OES • GFAAS	
As		0.5 mg/kg to 100 mg/kg	10 % to 2 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • GFAAS	
Ba		0.5 mg/kg to 100 mg/kg	10 % to 1 % (relative)	• ICP-MS • ID-ICP-MS	
Cd		0.01 mg/kg to 10 mg/kg	10 % to 2 % (relative)	• ICP-MS • ID-ICP-MS • ICP-OES • GFAAS	
Co		0.1 mg/kg to 10 mg/kg	10 % to 3 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • GFAAS	
Cr		0.1 mg/kg to 50 mg/kg	15 % to 2 % (relative)	• ID-ICP-MS • HR-ICP-MS • ICP-OES	
Cu		0.1 mg/kg to 50 mg/kg	10 % to 2 % (relative)	• ICP-MS • ID-ICP-MS • ICP-OES • GFAAS	
Fe		10 mg/kg to 1000 mg/kg	10 % to 2 % (relative)	• ICP-MS • ID-ICP-MS • ICP-OES • GFAAS	
Mn		0.1 mg/kg to 50 mg/kg	10 % to 2 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • GFAAS	
Ni		0.1 mg/kg to 10 mg/kg	15 % to 2 % (relative)	• ICP-MS • ID-ICP-MS • ICP-OES	
Pb		0.01 mg/kg to 10 mg/kg	15 % to 2 % (relative)	• ICP-MS • ID-ICP-MS • ICP-OES	
Zn		0.1 mg/kg to 100 mg/kg	10 % to 2 % (relative)	• ICP-MS • ID-ICP-MS • ICP-OES • GFAAS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method*1	Date of Accreditation
Food (trace elements and arsenic compounds in algae)	arsenate (As(V))	0.5 mg/kg to 100 mg/kg (as As)	10 % to 2 % (relative)	• HPLC-ICP-MS	2024-11-01
	arsenosugar-408 (arsenosugar-SO ₄)	0.1 mg/kg to 10 mg/kg (as As)	10 % to 2 % (relative)	• HPLC-ICP-MS	
	arsenosugar-328 (arsenosugar-OH)	0.1 mg/kg to 10 mg/kg (as As)	10 % to 2 % (relative)	• HPLC-ICP-MS	
	Hg	0.01 mg/kg to 0.1 mg/kg	10 % to 2 % (relative)	• ID-HR-ICP-MS	
Environmental matrix (trace elements in plant leaves)	Al	5 mg/kg to 5000 mg/kg	5 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • GFAAS	
	B	1 mg/kg to 500 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
	Ba	1 mg/kg to 500 mg/kg	10 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • ICP-OES	
	Ca	200 mg/kg to 20000 mg/kg	5 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS	
	Cd	0.005 mg/kg to 50 mg/kg	10 % to 3 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
	Co	0.01 mg/kg to 5 mg/kg	10 % to 2 % (relative)	• ICP-MS • HR-ICP-MS	
	Cu	0.5 mg/kg to 500 mg/kg	5 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • ICP-OES • GFAAS	
	Fe	0.5 mg/kg to 2000 mg/kg	10 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
	K	100 mg/kg to 30000 mg/kg	5 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS	
	Li	0.02 mg/kg to 10 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
	Mg	20 mg/kg to 5000 mg/kg	5 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS	
	Mn	5 mg/kg to 10000 mg/kg	5 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • GFAAS	
	Na	0.5 mg/kg to 100 mg/kg	20 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS	
	Ni	0.3 mg/kg to 100 mg/kg	10 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • ICP-OES	
	P	150 mg/kg to 10000 mg/kg	10 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES	
Pb	0.01 mg/kg to 100 mg/kg	20 % to 3 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS		
Rb	0.5 mg/kg to 200 mg/kg	10 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS		
Sr	0.5 mg/kg to 200 mg/kg	5 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • ICP-OES		
Zn	1 mg/kg to 500 mg/kg	10 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS		

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Food (trace elements in milk and dairy products)	Ca	0.5 g/kg to 100 g/kg	10 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS • FAES	2024-11-01
	Fe	0.01 g/kg to 10 g/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • ICP-OES	
	K	0.1 g/kg to 100 g/kg	10 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS • FAES	
	Mg	0.1 g/kg to 100 g/kg	10 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS	
	Na	0.01 g/kg to 50 g/kg	10 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES • FAAS • FAES	
	P	0.1 g/kg to 50 g/kg	10 % to 1 % (relative)	• ICP-MS • HR-ICP-MS • ICP-OES	
	Ba	0.05 mg/kg to 10 mg/kg	10 % to 1 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
	Cu	0.5 mg/kg to 100 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • GFAAS	
	Mn	0.1 mg/kg to 50 mg/kg	10 % to 2 % (relative)	• ICP-MS • HR-ICP-MS • GFAAS	
	Mo	0.02 mg/kg to 10 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
	Rb	0.1 mg/kg to 500 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
	Sr	0.1 mg/kg to 50 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS	
Zn	0.1 mg/kg to 1000 mg/kg	10 % to 2 % (relative)	• ID-ICP-MS • ICP-MS • HR-ICP-MS • ICP-OES		
High purity organic materials	creatinine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	2024-11-01
	urea	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	hydrocortisone	0.990 kg/kg to 1 kg/kg	0.001 kg/kg	• Subtracting method	
	isoleucine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	phenylalanine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	valine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	proline	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	alanine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	leucine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	lysine monohydrochloride	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	arginine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	uric acid	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	triolein	0.990 kg/kg to 1 kg/kg	0.001 kg/kg	• qNMR • Subtraction method	
	triglyceride	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• qNMR • Subtraction method	
	glycine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	glutamic acid	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	aspartic acid	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	tyrosine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	histidine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
	serine	0.990 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination	
threonine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination		
methionine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination		
cystine	0.995 kg/kg to 1 kg/kg	0.001 kg/kg	• Neutralization titration • Nitrogen determination		

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method * 1	Date of Accreditation
Organic standard solution	C-reactive protein	10 µmol/kg to 50 µmol/kg	2 % (relative)	• ID-LC-MS	2024-11-01
	total deoxyribonucleic acid (DNA) less than 650 bp	0.5 ng/µL to 200 ng/µL	5 % (relative)	• ID-LC-MS • ICP-MS	
	C-peptide	0.08 g/L to 1 g/L	3 % (relative)	• ID-LC-MS	
	total C-peptide (mixture of C-peptide, deamidated C-peptide, and pyroglutamylated C-peptide)	0.08 g/L to 1 g/L	3 % (relative)	• ID-LC-MS	
	total ribonucleic acid (RNA) less than 1100 bases	10 ng/µL to 200 ng/µL	4 % (relative)	• ID-LC-MS • ICP-MS	
	albumin	1 g/L to 100 g/L	1.6 % (relative)	• ID-LC-MS	
	okadaic acid	0.5 µg/mL to 10 µg/mL	4 % (relative)	• qNMR • Gravimetric preparation	
	dinophysistoxin-1	0.5 µg/mL to 10 µg/mL	1.6 % (relative)	• qNMR • Gravimetric preparation	
	monoclonal antibody	0.5 g/L to 100 g/L	2.6 % (relative)	• ID-LC-MS	
Environmental matrix (food)	okadaic acid	0.01 mg/kg to 10 mg/kg	10 % (relative)	• LC-MS	
	dinophysistoxin-1	0.01 mg/kg to 10 mg/kg	10 % (relative)	• LC-MS	
Steroids in serum	cortisol (hydrocortisone)	15 µg/L to 250 µg/L	3 % to 2 % (relative)	• ID-LC-MS	
	aldosterone	100 pg/mL to 1000 pg/mL	5 % (relative)	• ID-LC-MS	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method*1	Date of Accreditation
Molecular weight of polymer	poly (ethylene glycol) nonylphenyl ether (mass-average molecular mass, number-average molecular mass)	600 to 700	3 % (relative)	• SFC	2024-11-01
	poly (ethylene glycol) nonylphenyl ether (mass fraction and mole fraction of each degree of polymerization)	1×10^{-4} to 1	5 % (relative)	• SFC	
	polystyrene (mass-average molecular mass, number-average molecular mass, peak-average molecular mass)	400 to 2600	0.5 % (relative)	• SFC	
	polystyrene (polydispersity)	1.05 to 1.20	1.5 % (relative)	• SFC	
	polystyrene (mass fraction and mole fraction of each degree of polymerization)	2×10^{-5} to 1	2 % (relative)	• SFC	
	poly (ethylene glycol) (mass-average molecular mass, number-average molecular mass)	350 to 1700	1 % (relative)	• SFC	
	poly (ethylene glycol) (mass fraction and mole fraction of each degree of polymerization)	3×10^{-5} to 1	1 % (relative)	• SFC	
	monodisperse polystyrene (mass-average molar mass)	1×10^5 to 1×10^6	5 % (relative)	• Static light scattering (SLS)	
	poly (ethylene glycol) 23mer (mass fraction)	0.99 to 1	0.1 % (relative)	• SFC	
Particle reference material	polystyrene latex nanoparticle (light scattering intensity averaged diameter)	100 nm to 300nm	1 % (relative)	• Dynamic light scattering (DLS)	2024-11-01
Polymer reference material (polymer: organic compounds)	polybrominated diphenyl ether in plastics (polystyrene, polyvinyl chloride)	50 mg/kg to 1500 mg/kg	5 % to 2 % (relative)	• ID-GC-MS • HPLC	
	plasticizers (dimethyl phthalate, diethyl phthalate, di- <i>n</i> -propyl phthalate, di- <i>i</i> -butyl phthalate, di- <i>n</i> -butyl phthalate, di- <i>n</i> -pentyl phthalate, di- <i>n</i> -hexyl phthalate, dicyclohexyl phthalate, di- <i>n</i> -heptyl phthalate, butyl benzyl phthalate, bis(2-ethylhexyl) phthalate, bis(<i>n</i> -octyl) phthalate) in plastics (polystylen, polypropylene, polyvinyl chloride)	50 mg/kg to 1500 mg/kg	3 % to 1.5 % (relative)	• ID-GC-MS • HPLC	
Polymer reference material (Raman shift)	Raman shift	$300 \text{ cm}^{-1} \sim 3500 \text{ cm}^{-1}$	0.28 cm^{-1}	• Raman spectroscopy	
Polymer (perfluoroalkyl substances in polymer)	perfluorooctanesulfonic acid and its salts	10 mg/kg to 100 mg/kg	20 % to 10 % (relative)	• ID-LC-MS/MS	
Positron lifetime	positron lifetime in solids	0.1 ns to 20 ns	2 % (relative)	• Positron annihilation lifetime spectroscopy	
Steel	chromium	mass fraction 20 % to 40 %	0.1 % (relative)	• Titration • EPMA	
	nickel	mass fraction 15 % to 70 %	0.1 % (relative)	• Titration • EPMA	
	iron	mass fraction 5 % to 70 %	0.1 % (relative)	• Titration • EPMA	
	carbon	mass fraction 0.05 % to 1.0 %	10.0 % to 1.0 % (relative)	• Gravimetric analysis • EPMA	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method *1	Date of Accreditation
Thin film	film thickness	each layer 1 nm to 200 nm (total film thickness 3 nm to 200 nm or less)	0.27 % to 0.06 % (relative)	· X-ray reflectivity	2024-11-01
	arsenic	0.01 g/kg to 1.6 g/kg	2.4 % (relative)	· Instrumental Neutron Activation Analysis · ICP-MS	
Image sharpness evaluation	dot pitch	70 nm to 6000 nm	1.2 % (relative)	· SEM	
Thick film	film thickness	70 nm to 6000 nm	1.2 % (relative)	· SEM	

Subcategory	Measurand	Measurand Level or Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Analytical Method*1	Date of Accreditation
Thermophysical properties reference materials	Thermal expansion	-0.5×10 ⁻⁶ K ⁻¹ to 20×10 ⁻⁶ K ⁻¹ (Temperature range: 15 K to 1100 K)	0.005×10 ⁻⁶ K ⁻¹	· Laser interferometric thermal expansion measurement method	2024-11-01
	Thermal diffusivity	5×10 ⁻⁷ m ² s ⁻¹ to 2×10 ⁻⁴ m ² s ⁻¹ (Temperature range: 300 K to 1500 K)	3 % (relative)	· Laser flash method	
	Specific heat capacity	0.07 J K ⁻¹ g ⁻¹ to 1.8 J K ⁻¹ g ⁻¹ (Temperature range: 50 K to 900 K)	1 % (relative)	· Adiabatic calorimetry · Differential Scanning calorimetry	
	Thermal conductivity	1 W m ⁻¹ K ⁻¹ to 200 W m ⁻¹ K ⁻¹ (Temperature range: 300 K to 900 K)	5 % (relative)	The product of thermal diffusivity, specific heat capacity and density (thermal diffusivity: · laser flash method · pulse heating thermorefectance method specific heat capacity : · Adiabatic calorimetry · Differential Scanning calorimetry density: dimensions and weight)	
	Thermal diffusivity	3×10 ⁻⁶ m ² s ⁻¹ to 4×10 ⁻⁵ m ² s ⁻¹ (Measurement environment temperature: 5 °C to 35 °C)	6 % (relative)	· Pulse heating thermorefectance method	

*1

CRDS :	Cavity ring down spectroscopy
EPMA :	Electron probe microanalysis
DLS :	Dynamic light scattering
FAAS :	Flame atomic absorption spectrometry
FAES :	Flame atomic emission spectrometry
FI-ICP-MS :	Flow injection-inductively coupled plasma mass spectrometry
FT-IR :	Fourier transform infrared spectrometry
GC :	Gas chromatography
GC-ECD :	Gas chromatography/Electron capture detector
GC-FID :	Gas chromatography/Flame Ionization detector
GC-FPD :	Gas chromatography/Flame photometric detector
GC-MS :	Gas chromatography/Mass spectrometry
GC-PID :	Gas chromatography/Photo ionization detector
GC-SCD :	Gas chromatography/Sulfur chemiluminescence detector
GC-TCD :	Gas chromatography/Thermal conductivity detector
GFAAS :	Graphite furnace atomic absorption spectrometry
HPLC :	High performance liquid chromatography
HPLC-CAD	High performance liquid chromatography/Charged aerosol detector
HPLC-ICP-MS :	High performance liquid chromatography/inductively coupled plasma mass spectrometry
HPLC-UV	High performance liquid chromatography/Ultraviolet-visible absorption detector
HS-:	Head space-
HR-ICP-MS :	High-resolution inductively coupled plasma mass spectrometry
IC :	Ion chromatography
ICP-MS :	Inductively coupled plasma mass spectrometry
ICP-MS/MS :	Inductively coupled plasma tandem mass spectrometry
ICP-OES :	Inductively coupled plasma optical emission spectrometry
ID-GC-MS :	Isotope dilution-gas chromatography/mass spectrometry
ID-GC-ICP-MS :	Isotope dilution-gas chromatography/Inductively coupled plasma mass spectrometry
ID-HR-ICP-MS :	Isotope dilution-high-resolution inductively coupled plasma mass spectrometry
ID-HPLC-ICP-MS :	Isotope dilution-liquid chromatography/Inductively coupled plasma mass spectrometry
ID-ICP-MS :	Isotope dilution-inductively coupled plasma mass spectrometry
ID-ICP-MS/MS :	Isotope dilution-inductively coupled plasma tandem mass spectrometry
ID-LC-MS :	Isotope dilution-liquid chromatography/mass spectrometry
ID-LC-MS/MS :	Isotope dilution-liquid chromatography/tandem mass spectrometry
LC-MS :	Liquid chromatography/mass spectrometry
MC-ICP-MS :	Multicollector inductively coupled plasma mass spectrometry
MP-AES :	Microwave plasma atomic emission spectrometry
qNMR :	Quantitative nuclear magnetic resonance spectroscopy
SEM :	Scanning electron microscopy
SFC :	Supercritical fluid chromatography
SLS :	Static light scattering
TG :	Thermogravimetry

(End of Attachment)