



International Accreditation Japan

Information on Accredited Calibration Laboratory

Date of the update of the information : 2026-04-01

Accreditation Identification: ASNITE 0117 Calibration

Name of Calibration Laboratory : NTK COX CALIBRATION CENTER,
NTK INTERNATIONAL CORPORATION

Location of Calibration Laboratory: 417-1 Shimo-Ongata-machi, Hachioji-shi, Tokyo
192-0154, JAPAN

Name of Legal Entity: NTK INTERNATIONAL CORPORATION

Conformance Accreditation Standard: ISO/IEC 17025:2017

Expiry Date of Accreditation: 2028-10-01

Effective Date of Accreditation: 2024-10-02

General Field of Calibration: Fluid flow

Date of Initial Accreditation of the Field: 2014-07-10

Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facility

Calibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Calibration Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)
Liquid flow meters	Oil flow meters	From 0.0113 m ³ /h less than 0.0227 m ³ /h (From 0.05 GPM less than 0.1 GPM)	0.39 %
		From 0.0227 m ³ /h up to 70.41 m ³ /h (From 0.1 GPM up to 310 GPM)	0.37 %

#All Calibration Procedures are in-house procedures developed by this laboratory.

GPM: US Liquid Gallon per Minute

(End of Attachment)