



International Accreditation Japan

Information on Accredited Testing Laboratory

Date of the update of the information : 2026-04-01

Accreditation Identification: ASNITE 0069 Testing

Name of Testing Laboratory: ORGANO CORPORATION
R&D AND ENGINEERING, R&D CENTER

Location of Testing Laboratory: 4-4-1, Nishionuma, Minami-ku, Sagamihara-shi, Kanagawa
252-0332, JAPAN

Name of Legal Entity: ORGANO CORPORATION

Conformance Accreditation Standard: ISO/IEC 17025:2017

Expiry Date of Accreditation: 2028-10-22

Name of Laboratory: ORGANO CORPORATION
R&D AND ENGINEERING, R&D CENTER
Address of Laboratory: 4-4-1, Nishionuma, Minami-ku, Sagamihara-shi, Kanagawa 252-0332, JAPAN
Work to carry out: Control of management system, Service to the customer, Review of requests,
Sampling, Sample storage, Analytical test, Ensuring the validity of results,
Reporting of results

Accreditation Scope			Testing Items	Test Methods	Effective Date of Accreditation
Category	Sub-Category	Measurement Techniques			
Chemical Products	Water	ICP/MS	Na/Ultrapure Water (include sampling)	JIS K 0553 5.2* ¹	2024-10-23
			K/Ultrapure Water (include sampling)	JIS K 0553 6.2* ¹	
			Ca/Ultrapure Water (include sampling)	JIS K 0553 7.3* ¹	
			Mg/Ultrapure Water (include sampling)	JIS K 0553 8.3* ¹	
			Cu/Ultrapure Water (include sampling)	JIS K 0553 9.3* ¹	
			Zn/Ultrapure Water (include sampling)	JIS K 0553 10.3* ¹	
			Pb/Ultrapure Water (include sampling)	JIS K 0553 11.2* ¹	
			Cd/Ultrapure Water (include sampling)	JIS K 0553 12.3* ¹	
			Ni/Ultrapure Water (include sampling)	JIS K 0553 13.2* ¹	
			Co/Ultrapure Water (include sampling)	JIS K 0553 14.2* ¹	
			Mn/Ultrapure Water (include sampling)	JIS K 0553 15.3* ¹	
			Cr/Ultrapure Water (include sampling)	JIS K 0553 16.2* ¹	
			Al/Ultrapure Water (include sampling)	JIS K 0553 17.2* ¹	
			Fe/Ultrapure Water (include sampling)	JIS K 0553 18.3* ¹	

*1

JIS K 0553 4.3

Washing by an ultrasonic wave with filled by ultrapure water, instead of the method waiting for 16 hours with filled by Nitric acid (0.2 mol/L)

Accreditation Scope			Testing Items	Test Methods	Effective Date of Accreditation
Category	Sub-Category	Measurement Techniques			
Chemical Products	Water	TOC	TOC/ Ultrapure water, Raw water, Medical Manufacture water	JIS K 0551 4.4* ² JIS K 0102-1 19.2* ³ The 18 th Edition, The Japanese Pharmacopeia 2.59 EP11 2.2.44 USP43 643	2024-10-23

*2

JIS K 0551 4.4

Quantitative range: C 10 µg/L ~ 2000 µg/L

*3

JIS K 0102-1 19.2

Quantitative range: C 0.1 mg/L ~ 5 mg/L

(End of Attachment)