



International Accreditation Japan

## Information on Accredited Calibration Laboratory

Date of the update of the information : 2026-04-01

Accreditation Identification: **ASNITE 0143 Calibration**

Name of Calibration Laboratory : **Measurement & Calibration Center,  
Higashifuji Calibration & Measurement System Dept,  
Toyota Technical Development Corporation**

Location of Calibration Laboratory : **1200 Mishuku, Susono-shi, Sizuoka 410-1193, JAPAN**

Name of Legal Entity: **Toyota Technical Development Corporation**

Conformance Accreditation Standard: **ISO/IEC 17025:2017**

Expiry Date of Accreditation : **2028-11-10**

Effective Date of Accreditation: 2024-11-11

General Field of Calibration: Angular VelocityDate of Initial Accreditation of the Field: 2024-11-11Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facilityCalibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Calibration Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)
Angular Velocity Measuring Equipment	Inertial Measurement Unit	More than 0 rad/s up to $\frac{25\pi}{180}$ rad/s (More than 0°/s up to 25°/s)	$\frac{0.010\pi}{180}$ rad/s (0.010°/s)

#All Calibration Procedures are in-house procedures developed by this laboratory.

General Field of Calibration: AccelerationDate of Initial Accreditation of the Field: 2024-11-11Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facilityCalibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Calibration Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)
Static Acceleration Measuring Equipment, etc.	Inertial Measurement Unit	From -9.797 m/s <sup>2</sup> up to -0.100 m/s <sup>2</sup> From 0.100 m/s <sup>2</sup> up to 9.797 m/s <sup>2</sup>	0.006 m/s <sup>2</sup>

#All Calibration Procedures are in-house procedures developed by this laboratory.

*(End of Attachment)*