



International Accreditation Japan

Information on Accredited Testing Laboratory

Date of the update of the information : 2026-04-01

Accreditation Identification: ASNITE 0104 Testing

Name of Testing Laboratory: Soka Laboratory, Matsuyama Laboratory and
NishinihonLaboratory,
The Industrial Analysis Service Ltd.

Location of Testing Laboratory: (Soka Laboratory)
2-11-7, Yatsuka, Soka-shi, Saitama 340-0028, JAPAN
(Matsuyama Laboratory)
1, Matsuyama-cho, Moka-shi, Tochigi 321-4346, JAPAN
(Nishinihon Laboratory)
101-1, Nakaku Sakamoto aza doibata, Taka-cho, Taka gun, Hyogo 679-1132, JAPAN

Name of Legal Entity: The Industrial Analysis Service Ltd.

Conformance Accreditation Standard: ISO/IEC 17025:2017

Expiry Date of Accreditation : 2028-04-23

Name of Laboratory: Soka Laboratory, The Industrial Analysis Service Ltd.
 Address of Laboratory: 2-11-7, Yatsuka, Soka-shi, Saitama 340-0028, JAPAN
 Work to carry out: Control of management system, Analytical test, Reporting of results
 Date of Initial Accreditation for the Laboratory: 2013-12-20

Accreditation Scope			Testing Items	Test Methods	Effective Date of Accreditation
Category	Sub-Category	Measurement Techniques			
Chemical Products	Molding Articles and Components	Absorptiometry	Cr(VI)/ Metals	IEC 62321-7-1:2015	2024-04-24
			Cr(VI)/ Polymers and Electronics	IEC 62321-7-2:2017	2024-04-24
		ICP/MS	Cr, Cd, Hg, Pb/ Polymers, Metals and Electronics	IEC 62321-4:2017 IEC 62321-5:2013	2024-04-24
		IC	Cl, Br/ Fluororesin and Fluororubber *1	BS EN 14582:2016 *2	2024-04-24
			F, Cl, Br/ Polymers and Electronics	IEC 62321-3-2:2020	2024-04-24
			I/ Polymers and Electronics	IEC 62321-3-2:2020 Appendix D	2024-04-24
			F, Cl, Br, I/ Solders	JEITA ET 7304A Appendix B 2010	2024-04-24
		GC/MS	PBB, PBDE/ Polymers and Electronics	IEC 62321-6:2015	2024-04-24
			Phthalates (DEHP, BBP, DBP, DIBP) *3/ Polymers and Electronics	IEC 62321-8:2017 *4	2024-04-24
		Polymer	LC/MS	PFHxS, PFOS, PFHxA, PFOA, PFNA, PFDA, PFUnDA, PFDoDA, PFTTrDA, PFTDA *5/ Polymers	DIN CEN/TS 15968; DIN SPEC 1038:2010-11 EN 17681-1:2022

[NOTE]

- *1: The testing items are resin and rubber-related products that contain halogen compounds (Cl, Br), and do not contain components that do not gasify even when incinerated using an automatic combustion device (quartz tube combustion method).
- *2: Automatic combustion equipment is used instead of the oxygen bomb pre-incineration procedure specified in the BS EN 14582 standard.
- *3: DEHP: Di(2-ethylhexyl) phthalate, BBP: Butyl benzyl phthalate, DBP: Dibutyl phthalate, DIBP: Diisobutyl phthalate
- *4: Pyrolysis/Thermal Desorption-Gas Chromatography-Mass Spectrometry (Py/TD-GC-MS) is excluded.
- *5: PFHxS: Perfluorohexanesulfoic acid, PFOS: Perfluorooctanesulfoic acid, PFHxA: Perfluorohexanoic acid, PFOA: Perfluorooctanoic acid, PFNA: Perfluorononanoic acid, PFDA: Perfluorodecanoic acid, PFUnDA: Perfluoroundecanoic acid, PFDoDA: Perfluorododecanoic acid, PFTTrDA: Perfluorotridecanoic acid, PFTDA: Perfluorotetradecanoic acid

Name of Laboratory: Matsuyama Laboratory, The Industrial Analysis Service Ltd.
 Address of Laboratory: 1, Matsuyama-cho, Moka-shi, Tochigi 321-4346, JAPAN
 Work to carry out: Control of management system, Analytical test, Reporting of results
 Date of Initial Accreditation for the Laboratory: 2013-12-20

Accreditation Scope			Testing Items	Test Methods	Effective Date of Accreditation
Category	Sub-Category	Measurement Techniques			
Chemical Products	Metal	Optical Emission (Except for ICP/AES)	Al, Fe, Ni, Cu, Zn, Ge, As, Ag, Cd, In, Sn, Sb, Au, Pb, Bi/Solders	JIS Z 3910 14 *1	2024-04-24
		ICP/AES	Al, Fe, Ni, Cu, Zn, Ge, As, Ag, Cd, In, Sb, Au, Pb, Bi/Solders	JIS Z 3910 13 *2	2024-04-24
		ICP/MS	Al, Fe, Ni, Cu, Zn, Ge, As, Ag, Cd, In, Sb, Au, Pb, Bi/Solders	JIS Z 3910 13 *3	2024-04-24
		Titrimetry (Volumetric)	Ag / Solders	JIS Z 3910 9 *4	2024-04-24

[NOTE]

- *1 JIS Z 3910 14 The ICP analysis method is used to determine the concentration of the reference materials used in spark discharge atomic emission spectroscopy, but the measurement procedure is different from the method specified by JIS.
- *2 JIS Z 3910 13 Instead of aqua regia specified by JIS, a solution with a different mixing ratio of hydrochloric acid and nitric acid is used.
- *3 JIS Z 3910 13 ICP/MS and ICP/MS/MS are used instead of ICP/AES specified by JIS.
- *4 JIS Z 3910 9 Analyze using potentiometric titration method instead of potassium thiocyanate titration method specified by JIS.

Name of Laboratory: Nishinohon Laboratory, The Industrial Analysis Service Ltd.
 Address of Laboratory: 101-1, Nakaku Sakamoto aza doibata, Taka-cho, Taka-gun, Hyogo 679-1132, JAPAN
 Work to carry out: Control of management system, Analytical test, Reporting of results
 Date of Initial Accreditation for the Laboratory: 2024-04-24

Accreditation Scope			Testing Items	Test Methods	Effective Date of Accreditation
Category	Sub-Category	Measurement Techniques			
Chemical Products	Metal	Optical Emission (Except for ICP/AES)	Al, Fe, Ni, Cu, Zn, Ge, As, Ag, Cd, In, Sn, Sb, Au, Pb, Bi/Solders	JIS Z 3910 14 *1	2024-04-24

[NOTE]

- *1 JIS Z 3910 14 The ICP analysis method is used to determine the concentration of the reference materials used in spark discharge atomic emission spectroscopy, but the measurement procedure is different from the method specified by JIS.

(End of Attachment)